




Part 1 Liter Vessel
 Part Number 14143-11
 Serial Number P35 001685 - P35 001845



Test Certificate

| | | | USP / Ph.Eur. Tolerances | |
|-------------------|--------------------|---------------------------|--------------------------|--------------|
| Dimensions | Vessel | Inside Ø a | 102.9 to 105.1 mm | 98 – 106 mm |
| | | Height b | 200 mm | 160 – 210 mm |
| | | With hemispherical bottom | | |
| Material | Borosilicate glass | | | |

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopeia requirements (at date of approval, see below).

Navi Mumbai, 2022-08-24

24-08-2022 15:20:22

Moreshwar Pulekar
 SOTAX Quality Management
 C-64 | First Floor | TTC Industrial area | Turbhe | Navi Mumbai 400705 | India